

Certificate of Analysis



Client:
Orbitrex Peptides

Laboratory:
TrustPointe Analytics LLC
1743 142nd Ave Suite 4
Dorr, MI 49323



Verify at TrustPointeLIMS.com
Key=QH0XD8QN1B0D

Sample Information

Sample Name	SS-31 10mg 1 Vial Endo	Client SID	SS1125a	Published	2025-11-06 13:46
--------------------	------------------------	-------------------	---------	------------------	------------------

Results

Sample ID	SPL-2182
USP <85> Bacterial Endotoxin	<5.00 EU/mL
USP <85> Sample CV %	0.0 %
USP <85> Spike CV %	2.0 %
USP <85> Spike Recovery	110 %

Comments for SPL-2182

General

Original dilution 2mL; USP limit 5 EU/kg/hr

Approvals

Created by

Ashlee Lust
President

Reviewed by

Joshua Lust
Managing Director